



THE SOUTH INDIA TEXTILE RESEARCH ASSOCIATION

CENTRE OF EXCELLENCE FOR MEDICAL TEXTILES

Physical, Chemical & Biological Testing Laboratories

(ISO/IEC 17025 NABL ACCREDITED)

13/37, Avinashi Road, Aerodrome Post, Coimbatore - 641 014, INDIA

TEST REPORT

XXVII / A / E / 12002 / 20

To

Clique Medical Solutions Pvt. Ltd

B/1806, Dheeraj Dreams Building No 2

Plot No 642, LBS Marg, Bhandup West

Mumbai, Maharashtra, India

Pincode : 400078

Contact No(s) : 9892120574

Test Report No : **O2000460**

Sample Received On : **24-07-2020**

Inward Date : **24-07-2020**

Report Date: **25-07-2020**

Dear Sir / Madam,

This has reference to the sample(s) submitted by you for testing vide your letter reference no : **DT 17.07.2020.**

The results pertaining to your sample(s) are enclosed herewith.

Encl : Bill

This is a computer generated report, hence does not require signature.

IMPORTANT

This report is strictly CONFIDENTIAL. Its use for publicity, arbitration or as evidence in legal disputes is forbidden. Reference of sample(s) given by the party. Samples are not drawn by the laboratory. The above results are related to the samples tested. The report shall not be reproduced except in full, without the written approval of the laboratory. Authenticity of the report may be verified from our website www.lab.sitraonline.org.



THE SOUTH INDIA TEXTILE RESEARCH ASSOCIATION

CENTRE OF EXCELLENCE FOR MEDICAL TEXTILES

Physical, Chemical & Biological Testing Laboratories

(ISO/IEC 17025 NABL ACCREDITED)

13/37, Avinashi Road, Aerodrome Post, Coimbatore - 641 014, INDIA

Test Report No : O2000460

Report Date : 25-07-2020

Clique Medical Solutions Pvt. Ltd

Ref : DT 17.07.2020

Breathability - Differential pressure-IS 16289 <i>IS 16289:2014 Annexure C</i>	O2000460-1 <u>Sample Particulars</u> : Thermaissance Antibacterial Reusable Face Mask
Pa/cm ²	26.0

- End of Report -

Authorized Signatory

(Mechanical)

Report for Thermaissance,
a brand of Clique Medical
Solutions Pvt Ltd